

**THE UNIVERSITY OF MISSISSIPPI
LABORATORY SERVICES**

Phone (662) 915-5433 FAX (662) 915-5480

INSTRUMENT CALIBRATION REPORT

INSTRUMENT TYPE :		SERIAL NUMBER :	
----------------------	--	--------------------	--

SCALE RANGE	CALIBRATION SOURCE	DISTANCE (METERS)	CALCULATED READING	OBSERVED READING	NOTES

	Instrument reads within + or - 10% full scale on all scales.
	Instrument does not reads within + or - 10% full scale on all scales; all scales read within + or - 20% (graph required as noted).
	Some scale ranges could not be brought into calibration with + or - 20% as noted.
	Batteries replaced at time of calibration.
	Notes :

CALIBRATED BY :	DATE :
-----------------	--------